

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-23 °C	-23.465 °C	-24.0 °C	-0.535 °C	± 1.0 °C	± 4.0E-01 °C
2 °C	1.805 °C	1.8 °C	-0.005 °C	± 1.0 °C	± 4.0E-01 °C
4 °C	3.820 °C	3.8 °C	-0.020 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	7.871 °C	8.0 °C	0.129 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.039 °C	25.1 °C	0.061 °C	± 1.0 °C	± 4.0E-01 °C
32 %HR	31.909 %HR	27 %HR	-4.9 %HR	± 5 %HR	± 1.9E+00 %HR
56 %HR	55.570 %HR	51 %HR	-4.6 %HR	± 5 %HR	± 1.9E+00 %HR
80 %HR	80.049 %HR	83 %HR	3.0 %HR	± 5 %HR	± 1.9E+00 %HR