

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.653 °C	16.3 °C	0.647 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.2 °C	0.151 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.348 °C	50.2 °C	-0.148 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	24.946 %HR	20.0 %HR	-4.946 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	48.0 %HR	-2.550 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	79.0 %HR	3.440 %HR	± 5.0 %HR	± 1.9E+00 %HR