

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0 mm	1.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
2.0 mm	2.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
3.0 mm	3.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
5.0 mm	5.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
10.0 mm	10.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
15.0 mm	15.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
20.0 mm	20.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
25.0 mm	25.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
40.0 mm	40.00 mm	0 µm	± 30 µm	±	5.8E+00 µm