

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.653 °C	15.4 °C	-0.253 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.4 °C	-0.649 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.348 °C	49.5 °C	-0.848 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	24.946 %HR	21.0 %HR	-3.946 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	49.0 %HR	-1.550 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	79.0 %HR	3.440 %HR	± 5.0 %HR	± 1.9E+00 %HR