

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
11.86 RPM	12 RPM	0.14 RPM	± 1 RPM	±	5.8E-01 RPM
30.08 RPM	30 RPM	-0.08 RPM	± 1 RPM	±	5.9E-01 RPM
60.06 RPM	60 RPM	-0.06 RPM	± 1 RPM	±	6.1E-01 RPM