

Certificado No.: 57201		Equipo (Instrument): Inclinometro			
Fecha de Calibración: 2025-05-23		Marca (Brand): Stec			
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
10.0000 °	10.0667 °	0.0667 °		±	0.58 °
20.0000 °	20.0333 °	0.0333 °		±	0.58 °
25.0000 °	25.0333 °	0.0333 °		±	0.58 °
30.0000 °	30.0333 °	0.0333 °		±	0.58 °
45.0000 °	45.0500 °	0.0500 °		±	0.58 °
60.0000 °	60.0333 °	0.0333 °		±	0.58 °
75.0000 °	75.0167 °	0.0167 °		±	0.58 °
85.0000 °	85.0500 °	0.0500 °		±	0.58 °
90.0000 °	90.0167 °	0.0167 °		±	0.58 °