

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.653 °C                  | 15.3 °C            | -0.353 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.049 °C                  | 24.1 °C            | -0.949 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 50.348 °C                  | 49.4 °C            | -0.948 °C              | ± 1.0 °C                      | ± 1.1E+00 °C                         |
| 25 %HR                     | 24.946 %HR                 | 21.0 %HR           | -3.946 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.550 %HR                 | 50.0 %HR           | -0.550 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.560 %HR                 | 80.0 %HR           | 4.440 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |