

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.653 °C	14.9 °C	-0.753 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.1 °C	-0.949 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.348 °C	49.4 °C	-0.948 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	24.946 %HR	20.0 %HR	-4.946 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	45.0 %HR	-5.550 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	70.0 %HR	-5.560 %HR	± 5.0 %HR	± 1.9E+00 %HR