

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.453 °C	29.8 °C	-0.653 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	31.357 %HR	34 %HR	2.643 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.343 %HR	52 %HR	2.657 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.476 %HR	74 %HR	-0.476 %HR	± 5.0 %HR	± 1.9E+00 %HR