

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.453 °C	30.3 °C	-0.153 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	31.357 %HR	28 %HR	-3.357 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.343 %HR	46 %HR	-3.343 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.476 %HR	70 %HR	-4.476 %HR	± 5.0 %HR	± 1.9E+00 %HR