

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.453 °C	30.6 °C	0.147 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	31.357 %HR	29 %HR	-2.357 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.343 %HR	46 %HR	-3.343 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.476 %HR	70 %HR	-4.476 %HR	± 5.0 %HR	± 1.9E+00 %HR