

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	1.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	1.25 mm	1.26 mm	10 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	1.5 mm	1.51 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2 mm	2.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	3 mm	3.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5 mm	5.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	6 mm	6.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	8 mm	8.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	9 mm	9.01 mm	10 µm	± 30 µm	± 6.1E+00 µm