

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.063 in	0.0005 in	± 0.002 in	±	5.8E-04 in
0.1000 in	0.101 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.1250 in	0.126 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.2000 in	0.201 in	0.0010 in	± 0.002 in	±	5.9E-04 in
0.2500 in	0.252 in	0.0020 in	± 0.002 in	±	6.0E-04 in
0.3000 in	0.302 in	0.0020 in	± 0.002 in	±	6.2E-04 in
0.5000 in	0.502 in	0.0020 in	± 0.002 in	±	6.8E-04 in
1.0000 in	1.002 in	0.0020 in	± 0.002 in	±	9.1E-04 in