

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.0 mm	1.01 mm	10 µm	± 8 µm	±	5.8E+00 µm
1.5 mm	1.51 mm	10 µm	± 8 µm	±	5.8E+00 µm
2.0 mm	2.00 mm	0 µm	± 8 µm	±	5.8E+00 µm
5.0 mm	5.01 mm	10 µm	± 8 µm	±	5.8E+00 µm
8.0 mm	8.00 mm	0 µm	± 8 µm	±	5.8E+00 µm
10.0 mm	10.00 mm	0 µm	± 8 µm	±	5.8E+00 µm
15.0 mm	15.01 mm	10 µm	± 8 µm	±	5.8E+00 µm
20.0 mm	20.02 mm	20 µm	± 8 µm	±	5.8E+00 µm
25.0 mm	25.01 mm	10 µm	± 8 µm	±	5.8E+00 µm
30.0 mm	30.00 mm	0 µm	± 8 µm	±	5.8E+00 µm
35.0 mm	35.01 mm	10 µm	± 8 µm	±	5.8E+00 µm