

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1.0 mm            | 1.01 mm   | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 1.5 mm            | 1.51 mm   | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 2.0 mm            | 2.00 mm   | 0 µm          | ± 25 µm           | ±                           | 5.8E+00 µm |
| 5.0 mm            | 5.01 mm   | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 8.0 mm            | 8.00 mm   | 0 µm          | ± 25 µm           | ±                           | 5.8E+00 µm |
| 10.0 mm           | 10.00 mm  | 0 µm          | ± 25 µm           | ±                           | 5.8E+00 µm |
| 15.0 mm           | 15.01 mm  | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 20.0 mm           | 20.02 mm  | 20 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 25.0 mm           | 25.01 mm  | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |
| 30.0 mm           | 30.00 mm  | 0 µm          | ± 25 µm           | ±                           | 5.8E+00 µm |
| 35.0 mm           | 35.01 mm  | 10 µm         | ± 25 µm           | ±                           | 5.8E+00 µm |