

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
5 °C	4.502 °C	4.6 °C	0.098 °C	± 0.4 °C	± 4.0E-01 °C
20 %HR	19.813 %HR	30.0 %HR	10.187 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	44.721 %HR	51.1 %HR	6.379 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	73.4 %HR	5.611 %HR	± 3.0 %HR	± 1.9E+00 %HR