

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 29.701 °C                  | 28.8 °C            | -0.901 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 20 %HR                     | 19.813 %HR                 | 30 %HR             | 10.187 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 44.721 %HR                 | 51 %HR             | 6.379 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 67.789 %HR                 | 73 %HR             | 5.611 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |