

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.701 °C	28.9 °C	-0.801 °C	± 0.4 °C	± 4.0E-01 °C
20 %HR	19.813 %HR	34.7 %HR	14.887 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	44.721 %HR	54.9 %HR	10.179 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	77.6 %HR	9.811 %HR	± 3.0 %HR	± 1.9E+00 %HR