

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	29.751 °C	31.1 °C	1.349 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.351 %HR	30 %HR	1.649 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.977 %HR	47 %HR	-2.977 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.685 %HR	67 %HR	-4.685 %HR	± 5.0 %HR	± 1.9E+00 %HR