

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.751 °C	29.8 °C	0.049 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.351 %HR	31 %HR	2.649 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.977 %HR	48 %HR	-1.977 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.685 %HR	73 %HR	1.315 %HR	± 5.0 %HR	± 1.9E+00 %HR