

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.892 °C	30.0 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	30.8 %HR	1.702 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.382 %HR	46.8 %HR	-1.582 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.738 %HR	69.8 %HR	-2.938 %HR	± 3.0 %HR	± 1.9E+00 %HR