

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.892 °C	30.1 °C	0.208 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	31.0 %HR	1.902 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.382 %HR	47.6 %HR	-0.782 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.738 %HR	70.1 %HR	-2.638 %HR	± 3.0 %HR	± 1.9E+00 %HR