

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 29.892 °C                  | 30.0 °C            | 0.108 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.098 %HR                 | 30.8 %HR           | 1.702 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 48.382 %HR                 | 47.5 %HR           | -0.882 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 72.738 %HR                 | 70.1 %HR           | -2.638 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |