

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 1.000 mm                 | 1.003 mm         | 3 µm                 | ± 10 µm                  | ±                                  | 5.8E-01 µm |
| 1.250 mm                 | 1.252 mm         | 2 µm                 | ± 10 µm                  | ±                                  | 5.8E-01 µm |
| 1.500 mm                 | 1.502 mm         | 2 µm                 | ± 10 µm                  | ±                                  | 5.8E-01 µm |
| 2.000 mm                 | 2.000 mm         | 0 µm                 | ± 10 µm                  | ±                                  | 5.8E-01 µm |
| 3.000 mm                 | 3.002 mm         | 2 µm                 | ± 10 µm                  | ±                                  | 5.8E-01 µm |
| 5.000 mm                 | 5.010 mm         | 10 µm                | ± 10 µm                  | ±                                  | 5.9E-01 µm |
| 8.000 mm                 | 8.012 mm         | 12 µm                | ± 10 µm                  | ±                                  | 5.9E-01 µm |
| 10.000 mm                | 10.015 mm        | 15 µm                | ± 10 µm                  | ±                                  | 5.9E-01 µm |
| 12.000 mm                | 12.019 mm        | 19 µm                | ± 10 µm                  | ±                                  | 5.9E-01 µm |