

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.020 °C	-0.2 °C	-0.180 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.361 °C	40.4 °C	0.039 °C	± 0.5 °C	± 4.0E-01 °C