

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.892 °C	29.8 °C	-0.092 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.098 %HR	33.0 %HR	3.902 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.382 %HR	50.5 %HR	2.118 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.738 %HR	71.2 %HR	-1.538 %HR	± 3.0 %HR	± 1.9E+00 %HR