

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.05	0.04919	0.810	± 1.2	± 4.3E-01 mm					
0 a 0.1	0.09921	0.790	± 1.2	± 4.3E-01 mm					
0 a 0.15	0.14918	0.820	± 1.3	± 4.3E-01 mm					
0 a 1	0.99923	0.770	± 2	± 4.3E-01 mm					
0 a 2	1.99928	0.720	± 2.8	± 4.8E-01 mm					
0 a 3	2.99932	0.680	± 3.6	± 4.8E-01 mm					
0 a 4	3.99936	0.640	± 4.4	± 4.8E-01 mm					
0 a 5	4.99941	0.590	± 5.2	± 4.8E-01 mm					
0 a 6	5.99946	0.540	± 6	± 3.6E-01 mm					
0 a 7	6.99950	0.500	± 6.8	± 3.6E-01 mm					
0 a 8	7.99954	0.460	± 7.6	± 3.6E-01 mm					
0 a 9	8.99958	0.420	± 8.4	± 3.6E-01 mm					
0 a 10	9.99963	0.370	± 9.2	± 3.6E-01 mm					
0 a 11	10.99968	0.320	± 10	± 3.6E-01 mm					
0 a 12	11.99972	0.280	± 10.8	± 3.6E-01 mm					
0 a 13	12.99975	0.250	± 11.6	± 3.6E-01 mm					
0 a 14	13.99980	0.200	± 12.4	± 3.6E-01 mm					
0 a 15	14.99984	0.160	± 13.2	± 3.6E-01 mm					