

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.481 °C	0.7 °C	0.219 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	14.6 °C	-0.421 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.4 °C	0.338 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	34.8 %HR	4.230 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	52.0 %HR	1.880 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	72.3 %HR	1.852 %HR	± 5.0 %HR	± 1.9E+00 %HR