

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 0.05	0.05087	-0.010	± 0.016	± 2.9E-01 in					
0 a 0.1	0.10096	-0.012	± 0.2	± 2.9E-01 in					
0 a 0.15	0.15096	-0.012	± 0.2	± 2.9E-01 in					
0 a 1	1.00093	-0.011	± 0.4	± 2.9E-01 in					
0 a 2	2.00087	-0.010	± 0.6	± 2.9E-01 in					
0 a 3	3.00008	-0.001	± 0.8	± 2.9E-01 in					
0 a 4	4.00013	-0.002	± 1	± 2.9E-01 in					
0 a 5	5.00018	-0.002	± 1.2	± 2.9E-01 in					
0 a 8	8.00074	-0.009	± 1.8	± 2.9E-01 in					
0 a 10	10.00077	-0.009	± 2.2	± 2.9E-01 in					
0 a 12	12.00088	-0.011	± 2.6	± 2.9E-01 in					
0 a 15	15.00020	-0.002	± 3.2	± 2.9E-01 in					
0 a 20	20.00009	-0.001	± 4.2	± 2.9E-01 in					
0 a 25	25.00092	-0.011	± 5.2	± 2.9E-01 in					
0 a 30	30.00097	-0.012	± 6.2	± 2.9E-01 in					