

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.05	0.04957	0.430	± 0.6	±	4.3E-01 mm						
0 a 0.1	0.09966	0.340	± 0.6	±	4.3E-01 mm						
0 a 0.15	0.14967	0.330	± 0.6	±	4.3E-01 mm						
0 a 1	0.99967	0.330	± 1	±	4.3E-01 mm						
0 a 2	1.99966	0.340	± 1.4	±	4.8E-01 mm						
0 a 3	2.99911	0.890	± 1.8	±	4.8E-01 mm						
0 a 4	3.99910	0.900	± 2.2	±	4.8E-01 mm						
0 a 5	4.99909	0.910	± 2.6	±	4.8E-01 mm						
0 a 8	7.99977	0.230	± 3.8	±	3.6E-01 mm						
0 a 10	9.99988	0.120	± 4.6	±	3.6E-01 mm						
0 a 12	12.00007	-0.070	± 5.4	±	3.6E-01 mm						
0 a 15	14.99961	0.390	± 6.6	±	3.6E-01 mm						
0 a 20	19.99969	0.310	± 8.6	±	3.6E-01 mm						
0 a 25	25.00060	-0.600	± 10.6	±	5.5E-01 mm						
0 a 30	30.00081	-0.810	± 12.6	±	5.5E-01 mm						