

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04960	0.400		± 4.3E-01 mm					
0 a 0.1	0.09966	0.340		± 4.3E-01 mm					
0 a 0.15	0.14956	0.440		± 4.3E-01 mm					
0 a 1	0.99992	0.080		± 4.3E-01 mm					
0 a 2	1.99966	0.340		± 4.8E-01 mm					
0 a 3	2.99944	0.560		± 4.8E-01 mm					
0 a 4	3.99967	0.330		± 4.8E-01 mm					
0 a 5	4.99982	0.180		± 4.8E-01 mm					
0 a 8	7.99991	0.090		± 3.6E-01 mm					
0 a 10	9.99988	0.120		± 3.6E-01 mm					
0 a 12	12.00027	-0.270		± 3.6E-01 mm					
0 a 15	14.99971	0.290		± 3.6E-01 mm					
0 a 20	19.99979	0.210		± 3.6E-01 mm					
0 a 25	25.00023	-0.230		± 5.5E-01 mm					
0 a 30	30.00029	-0.290		± 5.5E-01 mm					