

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04921	0.790	± 1.2	± 4.3E-01 mm					
0 a 0.1	0.09948	0.520	± 1.2	± 4.3E-01 mm					
0 a 0.15	0.14963	0.370	± 1.3	± 4.3E-01 mm					
0 a 1	0.99935	0.650	± 2	± 4.3E-01 mm					
0 a 2	1.99934	0.660	± 2.8	± 4.8E-01 mm					
0 a 3	2.99924	0.760	± 3.6	± 4.8E-01 mm					
0 a 4	3.99926	0.740	± 4.4	± 4.8E-01 mm					
0 a 5	4.99935	0.650	± 5.2	± 4.8E-01 mm					
0 a 8	7.99944	0.560	± 7.6	± 3.6E-01 mm					
0 a 10	9.99952	0.480	± 9.2	± 3.6E-01 mm					
0 a 12	11.99953	0.470	± 10.8	± 3.6E-01 mm					
0 a 15	14.99958	0.420	± 13.2	± 3.6E-01 mm					
0 a 20	19.99990	0.100	± 17.2	± 3.6E-01 mm					
0 a 25	25.00011	-0.110	± 21.2	± 5.5E-01 mm					
0 a 30	30.00024	-0.240	± 25.2	± 5.5E-01 mm					