

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04986	0.140		± 4.3E-01 mm					
0 a 0.1	0.09969	0.310		± 4.3E-01 mm					
0 a 0.15	0.14971	0.290		± 4.3E-01 mm					
0 a 1	0.99992	0.080		± 4.3E-01 mm					
0 a 2	1.99993	0.070		± 4.8E-01 mm					
0 a 3	3.00012	-0.120		± 4.8E-01 mm					
0 a 4	3.99973	0.270		± 4.8E-01 mm					
0 a 5	4.99997	0.030		± 4.8E-01 mm					
0 a 8	8.00015	-0.150		± 3.6E-01 mm					
0 a 10	10.00017	-0.170		± 3.6E-01 mm					
0 a 12	12.00006	-0.060		± 3.6E-01 mm					
0 a 15	14.99967	0.330		± 3.6E-01 mm					
0 a 20	19.99990	0.100		± 3.6E-01 mm					
0 a 25	25.00014	-0.140		± 5.5E-01 mm					
0 a 30	30.00020	-0.200		± 5.5E-01 mm					