

Certificado No. (Certificate #):		57369		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-05-23		Marca (Brand):		JAGUAR	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		10 mm	10.00 mm	0 µm	± 50 µm	± 4.1E+01 µm	
Exterior / (Exterior)		20 mm	20.00 mm	0 µm	± 50 µm	± 4.1E+01 µm	
Exterior / (Exterior)		50 mm	50.00 mm	0 µm	± 50 µm	± 4.1E+01 µm	
Exterior / (Exterior)		100 mm	100.00 mm	0 µm	± 50 µm	± 4.1E+01 µm	
Exterior / (Exterior)		200 mm	200.10 mm	100 µm	± 70 µm	± 4.1E+01 µm	
Interior / (Interior)		50 mm	50.00 mm	0 µm	± 50 µm	± 3.1E+01 µm	
Interior / (Interior)		150 mm	150.05 mm	50 µm	± 50 µm	± 3.1E+01 µm	
Profundidad / (Depth)		30 mm	30.00 mm	0 µm	± 50 µm	± 3.3E+01 µm	
Profundidad / (Depth)		100 mm	100.00 mm	0 µm	± 50 µm	± 3.3E+01 µm	