

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	14.7 °C	-0.522 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.1 °C	-0.336 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.043 °C	23.1 °C	0.057 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	26.0 %HR	-4.366 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	45.8 %HR	-4.279 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	67.0 %HR	-3.345 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	71.0 %HR	-4.560 %HR	± 5.0 %HR	± 1.9E+00 %HR