

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.0 °C	-0.222 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.3 °C	-0.136 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.043 °C	22.9 °C	-0.143 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.2 °C	0.138 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	32.0 %HR	1.634 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	49.0 %HR	-1.079 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	72.0 %HR	1.655 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	77.0 %HR	1.440 %HR	± 5.0 %HR	± 1.9E+00 %HR