

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.0 °C	-0.222 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.3 °C	-0.136 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.043 °C	22.9 °C	-0.143 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	26.0 %HR	-4.366 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	46.0 %HR	-4.079 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	68.0 %HR	-2.345 %HR	± 5 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	72.0 %HR	-3.560 %HR	± 5 %HR	± 1.9E+00 %HR