

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.079 %HR	45.0 %HR	-5.079 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.304 %HR	72.0 %HR	1.696 %HR	± 5.0 %HR	± 1.9E+00 %HR