

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.2 °C	-0.022 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.6 °C	0.538 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.079 %HR	30.0 %HR	-20.079 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.304 %HR	57.0 %HR	-13.304 %HR	± 5.0 %HR	± 1.9E+00 %HR