

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.158 °C	10.1 °C	-0.058 °C	± 0.4 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 0.4 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.4 °C	± 4.0E-01 °C
32 °C	31.867 °C	31.6 °C	-0.267 °C	± 0.4 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.1 %HR	-1.470 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	66.7 %HR	-3.748 %HR	± 4.0 %HR	± 1.9E+00 %HR