

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.158 °C         | 9.9 °C    | -0.258 °C     | ± 0.4 °C             | ± 4.0E-01 °C                |
| 15 °C             | 15.021 °C         | 14.8 °C   | -0.221 °C     | ± 0.4 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.349 °C         | 25.0 °C   | -0.349 °C     | ± 0.4 °C             | ± 4.0E-01 °C                |
| 32 °C             | 31.867 °C         | 31.8 °C   | -0.067 °C     | ± 0.4 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 30.2 %HR  | -0.370 %HR    | ± 4.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 66.7 %HR  | -3.748 %HR    | ± 4.0 %HR            | ± 1.9E+00 %HR               |