

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.100 in	0.099 in	-0.001 in	± 0.001 in	±	5.8E-04 in
0.125 in	0.123 in	-0.002 in	± 0.001 in	±	5.8E-04 in
0.200 in	0.198 in	-0.002 in	± 0.001 in	±	5.9E-04 in
0.250 in	0.248 in	-0.002 in	± 0.001 in	±	6.0E-04 in
0.300 in	0.298 in	-0.002 in	± 0.001 in	±	6.1E-04 in
0.350 in	0.348 in	-0.002 in	± 0.001 in	±	6.3E-04 in
0.400 in	0.398 in	-0.002 in	± 0.001 in	±	6.4E-04 in
0.450 in	0.447 in	-0.003 in	± 0.001 in	±	6.6E-04 in
0.500 in	0.496 in	-0.004 in	± 0.001 in	±	6.7E-04 in