

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.032 °C         | 29.9 °C   | -0.132 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.366 %HR        | 41 %HR    | 10.534 %HR    | ± 2.5 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.499 %HR        | 56 %HR    | 5.401 %HR     | ± 2.5 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 77 %HR    | 6.552 %HR     | ± 2.5 %HR            | ± 1.9E+00 %HR               |