

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.065 in	0.0025 in	± 0.002 in	±	5.8E-04 in
0.1000 in	0.101 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.1250 in	0.126 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.2000 in	0.201 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.2500 in	0.251 in	0.0010 in	± 0.002 in	±	5.8E-04 in
0.5000 in	0.501 in	0.0010 in	± 0.002 in	±	5.8E-04 in
1.0000 in	1.001 in	0.0010 in	± 0.002 in	±	5.8E-04 in
2.0000 in	2.001 in	0.0010 in	± 0.003 in	±	5.8E-04 in
4.0000 in	4.001 in	0.0010 in	± 0.003 in	±	5.8E-04 in