

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 25 °C                      | 24.848 °C                  | 25.1 °C            | 0.252 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| # 25 °C                    | 24.848 °C                  | 25.2 °C            | 0.352 °C               | ± 0.5 °C                      | ± 7.4E-02 °C                         |
| # 50 °C                    | 49.917 °C                  | 50.3 °C            | 0.383 °C               | ± 0.5 °C                      | ± 7.4E-02 °C                         |
| # 70 °C                    | 69.972 °C                  | 70.1 °C            | 0.128 °C               | ± 0.5 °C                      | ± 7.4E-02 °C                         |
| 25 %HR                     | 23.413 %HR                 | 25.3 %HR           | 1.887 %HR              | ± 3 %HR                       | ± 1.9E+00 %HR                        |
| 50 %HR                     | 48.975 %HR                 | 50.5 %HR           | 1.525 %HR              | ± 3 %HR                       | ± 1.9E+00 %HR                        |
| 70 %HR                     | 69.425 %HR                 | 71.0 %HR           | 1.575 %HR              | ± 3 %HR                       | ± 1.9E+00 %HR                        |