

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
130.3 μm	130.0 μm	-0.3 μm	± 5 μm	±	5.8E-01 μm
523.0 μm	523.2 μm	0.2 μm	± 5 μm	±	5.8E-01 μm