

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.005 mm	1.008 mm	3 μm	± 50.25 μm	±	5.8E-01 μm
1.5 mm	1.503 mm	3 μm	± 75. μm	±	5.8E-01 μm
2 mm	2.004 mm	4 μm	± 100 μm	±	5.8E-01 μm
4 mm	4.003 mm	3 μm	± 200 μm	±	5.8E-01 μm
5 mm	5.001 mm	1 μm	± 250 μm	±	5.8E-01 μm