

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.161 °C	15.8 °C	0.639 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.343 °C	31 °C	0.357 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.325 %HR	27 %HR	-3.325 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.785 %HR	60 %HR	-10.785 %HR	± 5.0 %HR	± 1.9E+00 %HR