

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.161 °C	14.7 °C	-0.461 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.343 °C	29.6 °C	-0.743 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.325 %HR	30 %HR	-0.325 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.785 %HR	65 %HR	-5.785 %HR	± 5.0 %HR	± 1.9E+00 %HR