

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.809 °C	11.1 °C	0.291 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.870 °C	33.1 °C	0.230 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	34.8 %HR	-0.883 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	77.8 %HR	2.240 %HR	± 3.0 %HR	± 1.9E+00 %HR